Ref. Certif. No.

SI-5780

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST CERTIFICATES FOR ELECTRICAL EQUIPMENT (IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE CERTIFICATS D'ESSAIS DES EQUIPEMENTS ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product Produit	DIN Rail power supply for medical application
Name and address of the applicant Nom et adresse du demandeur	PULS GmbH Arabellastr. 15, D-81925 München, Germany
Name and address of the manufacturer Nom et adresse du fabricant	PULS GmbH Elektrastr. 6, D-81925 München, Germany
Name and address of the factory Nom et adresse de l'usine *Note: When more than one factory, please report on page 2* *Note: Lorsque il y plus d'une usine, veuillez utiliser la 2ème page*	See page 2 ☒ Additional Information on page 2
Ratings and principal characteristics Valeurs nominales et caractéristiques principales	See page 2
Trademark (if any) Marque de fabrique (si elle existe)	PULS
Type of Manufacturer's Testing Laboratories used Type de programme du laboratoire d'essais constructeur	/
Model / Type Ref. Ref. De type	CP10.241-M1 and CP10.241-M1-78
Additional information (if necessary may also be reported on page 2) Les informations complémentaires (si nécessaire, peuvent être indiqués sur la 2ème page	The risk management requirements of the standard were not addressed. ☒ Additional Information on page 2
A sample of the product was tested and found to be in conformity with Un échantillon de ce produit a été essayé et a été considéré conforme à la	IEC 60601-1:2005 (Third Edition) + A1:2012
As shown in the Test Report Ref. No. which forms part of this Certificate Comme indiqué dans le Rapport d'essais numéro de référence qui constitue partie de ce Certificat	T221-0097/16 issued by laboratory SIQ Testing and Certification GmbH

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme **National de Certification**

SIQ Ljubljana
Tržaška cesta 2, SI-1000 Ljubljana, Slovenia
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si

Date: 2016-12-20

Signature: Bojan Pečavar

LRI Distribuída no Brasil pela LRI | vendas@lri.com.br | 51 2129-3000

Ratings:
Input:
AC: 100-240 Vac; 3,3-1,4 A; 50-60 Hz
DC: 110-150 Vdc; 3,0-2,2 A

CP10.241-M1 output:
24-28 Vdc; 12-10,3 A (below +45°C)
24-28 Vdc; 10-8,6 A (below +60°C)
24-28 Vdc; 7,5-6,5 A (below +70°C)

CP10.241-M1-78 output:
24-28 Vdc; 10-8,6 A (below +60°C)
24-28 Vdc; 7,5-6,5 A (below +70°C)

Name and address of factories:

1) PULS ELECTRONICS (SUZHOU) CO LTD
NO 1 RUI EN LANE XINGPU RD, SIP
SUZHOU
JIANGSU CN-215021, CHINA

2) PULS INVESTICNI S R O
PRAZSKA 5639
CZ-430 01 CHOMUTOV, CZECH REPUBLIC

Additional information (if necessary)
Information complémentaire (si nécessaire)

Date: 2016-12-20

Signature: Bojan Pečavar

Distribuída no Brasil pela LRI | vendas@lri.com.br | 51 2129-3000

SLOVENSKA AKREDITACIJA
SIST EN ISO/IEC 17065
CP-001

Certificate of Conformity

Number:	**C251-0129/16**	Project file:	C20161277

Product: **DIN-rail power supply for medical application**

Type reference: **CP10.241-M1**

Trademark: PULS

Applicant: **PULS GmbH**
Arabellastrasse 15, DE-81925 München, Germany

Manufacturer: PULS GmbH
Arabellastrasse 15, DE-81925 München, Germany

Place of manufacture: PULS GmbH
Arabellastrasse 15, DE-81925 München, Germany

This certificate is granted subject to the SIQ's rules on product certification. SIQ certifies the conformity of the products with the requirements of the listed standards.

Ratings: AC input: 100 - 240 V~; 50 - 60 Hz; 288 W max (at < 45°C)
DC input: 110 - 150 Vdc
DC output: adjustment range 24 – 28 Vdc / 12 – 10.3 A (at < 45°C)

Standard: EN 60601-1-2:2015

Test report: T251-0895/16 (2016-11-04)

Remarks: This certificate shall apply to the products identical to the tested sample and shall remain valid for the period of 3 years until 2019-12-21 or until the validity date of the listed standards, whichever occurs earlier.

Date: 2016-12-21 **Authorized signature:** Bojan Pečavar

LRI Distribuída no Brasil pela LRI | vendas@lri.com.br | 51 2129-3000

Only integral publication of this certificate is allowed. This certificate may only be reproduced in its entirety and without any changes. On request SIQ will give information about the validity of the certificate.

SIQ Ljubljana, Tržaška cesta 2, SI-1000 Ljubljana, Slovenia
T +386 1 4778 100, F +386 1 4778 444, info@siq.si, www.siq.si